



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031975.D
Acq On : 07 Oct 2022 21:41
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Title : SW846 8260
Last Update : Thu Oct 06 14:28:46 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	5193	PASS
75	95	30	60	52.9	14664	PASS
95	95	100	100	100.0	27728	PASS
96	95	5	9	7.5	2081	PASS
173	174	0.00	2	1.2	255	PASS
174	95	50	100	76.6	21227	PASS
175	174	5	9	7.4	1576	PASS
176	174	95	101	96.3	20449	PASS
177	176	5	9	6.8	1385	PASS