

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100724\
 Data File : VX043304.D
 Acq On : 07 Oct 2024 11:07
 Operator : JC/MD
 Sample : P4300-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-5

Quant Time: Oct 08 01:43:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87476	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	163874	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72966	51.075	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.140%	
35) Dibromofluoromethane	5.379	113	53762	47.571	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.140%	
50) Toluene-d8	8.647	98	193774	49.477	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.960%	
62) 4-Bromofluorobenzene	11.079	95	67696	47.579	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 95.160%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100724\
Data File : VX043304.D
Acq On : 07 Oct 2024 11:07
Operator : JC/MD
Sample : P4300-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-5

Quant Time: Oct 08 01:43:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

