

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100724\
 Data File : VX043307.D
 Acq On : 07 Oct 2024 12:16
 Operator : JC/MD
 Sample : P4263-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20240930DL

Quant Time: Oct 08 01:44:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	160451	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139774	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73276	52.643	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.280%			
35) Dibromofluoromethane	5.385	113	53776	48.599	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.200%			
50) Toluene-d8	8.647	98	191159	49.850	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.700%			
62) 4-Bromofluorobenzene	11.079	95	67405	48.385	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.760%			
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	2354	2.469	ug/l	97
12) 1,1-Dichloroethene	2.313	96	849	0.897	ug/l	91
44) Trichloroethene	7.123	130	49188	43.756	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100724\
Data File : VX043307.D
Acq On : 07 Oct 2024 12:16
Operator : JC/MD
Sample : P4263-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20240930DL

Quant Time: Oct 08 01:44:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

