

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100724\
 Data File : VX043308.D
 Acq On : 07 Oct 2024 12:39
 Operator : JC/MD
 Sample : P4285-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WATER-TOTESDL

Quant Time: Oct 08 01:45:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	87526	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	165293	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	143778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74315	51.989	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.980%
35) Dibromofluoromethane	5.385	113	54160	47.512	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.020%
50) Toluene-d8	8.647	98	197876	50.091	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.180%
62) 4-Bromofluorobenzene	11.079	95	70861	49.376	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.760%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.989	59	8257	30.355	ug/l	98
16) Acetone	2.386	43	57456	96.250	ug/l	97
25) 2-Butanone	4.562	43	70813	86.517	ug/l	98
29) Tetrahydrofuran	5.007	42	48772	93.538	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100724\
Data File : VX043308.D
Acq On : 07 Oct 2024 12:39
Operator : JC/MD
Sample : P4285-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WATER-TOTESDL

Quant Time: Oct 08 01:45:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

