

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100725\
 Data File : VX048044.D
 Acq On : 07 Oct 2025 11:59
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 08 01:25:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	143940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	293410	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	308738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	160937	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	127141	55.492	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.980%
35) Dibromofluoromethane	5.373	113	100790	51.220	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.440%
50) Toluene-d8	8.634	98	340519	50.011	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.020%
62) 4-Bromofluorobenzene	11.067	95	151399	58.589	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100725\
Data File : VX048044.D
Acq On : 07 Oct 2025 11:59
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 08 01:25:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

