

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100725\
 Data File : VX048047.D
 Acq On : 07 Oct 2025 13:02
 Operator : JC/MD
 Sample : Q3298-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-TB-20251003

Quant Time: Oct 08 01:27:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	119485	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	242388	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	248619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	121141	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	96225	50.594	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.180%
35) Dibromofluoromethane	5.373	113	79183	48.710	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.420%
50) Toluene-d8	8.635	98	268662	47.763	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.067	95	115010	53.876	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100725\
Data File : VX048047.D
Acq On : 07 Oct 2025 13:02
Operator : JC/MD
Sample : Q3298-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-TB-20251003

Quant Time: Oct 08 01:27:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

