

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_X\DATA\VX100818\  
 Data File : VX005113.D  
 Acq On : 08 Oct 2018 08:49  
 Operator : JC/MD  
 Sample : VX1008WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX1008WBL01

Quant Time: Oct 08 17:13:06 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X092618W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 03 04:00:03 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 5.67  | 168  | 210842   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.86  | 114  | 332860   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.12 | 117  | 322002   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.08 | 152  | 175384   | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds |        |     |          |       |         |      |
|-----------------------------|--------|-----|----------|-------|---------|------|
| 33) 1,2-Dichloroethane-d4   | 6.07   | 65  | 154930   | 51.44 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 102.88% |      |
| 35) Dibromofluoromethane    | 5.50   | 113 | 132092   | 46.82 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 93.64%  |      |
| 50) Toluene-d8              | 8.71   | 98  | 474117   | 50.55 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 101.10% |      |
| 62) 4-Bromofluorobenzene    | 11.14  | 95  | 168995   | 50.02 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 100.04% |      |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA\_X\DATA\VX100818\  
Data File : VX005113.D  
Acq On : 08 Oct 2018 08:49  
Operator : JC/MD  
Sample : VX1008WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1008WBL01

Quant Time: Oct 08 17:13:06 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X092618W.M  
Quant Title : SW846 8260  
QLast Update : Wed Oct 03 04:00:03 2018  
Response via : Initial Calibration

