

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
 Data File : VX005118.D
 Acq On : 08 Oct 2018 11:07
 Operator : JC/MD
 Sample : J5304-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CA-AQIDW-20181004

Quant Time: Oct 09 12:47:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	198444	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	312189	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	307586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	173771	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.07	65	147580	52.06	ug/l	0.00
Spiked Amount	50.000			Recovery	=	104.12%	
35) Dibromofluoromethane		5.51	113	125556	47.45	ug/l	0.00
Spiked Amount	50.000			Recovery	=	94.90%	
50) Toluene-d8		8.72	98	448363	50.97	ug/l	0.00
Spiked Amount	50.000			Recovery	=	101.94%	
62) 4-Bromofluorobenzene		11.14	95	163013	51.44	ug/l	0.00
Spiked Amount	50.000			Recovery	=	102.88%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
Data File : VX005118.D
Acq On : 08 Oct 2018 11:07
Operator : JC/MD
Sample : J5304-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CA-AQIDW-20181004

Quant Time: Oct 09 12:47:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

