

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
 Data File : VX005123.D
 Acq On : 08 Oct 2018 13:21
 Operator : JC/MD
 Sample : J5189-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-131D3-20180927

Quant Time: Oct 09 12:50:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	189607	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	301912	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	298746	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	169228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	143505	52.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.96%	
35) Dibromofluoromethane	5.50	113	123360	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
50) Toluene-d8	8.72	98	437939	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
62) 4-Bromofluorobenzene	11.14	95	157788	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	369293	178.02	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	1738	0.69	ug/l	76
44) Trichloroethene	7.21	130	27309	10.16	ug/l	96
64) Tetrachloroethene	9.34	164	4357	1.48	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
Data File : VX005123.D
Acq On : 08 Oct 2018 13:21
Operator : JC/MD
Sample : J5189-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-131D3-20180927

Quant Time: Oct 09 12:50:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

