

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_X\DATA\VX100818\  
 Data File : VX005124.D  
 Acq On : 08 Oct 2018 13:48  
 Operator : JC/MD  
 Sample : J5189-07  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE-105D1-20180927

Quant Time: Oct 09 12:53:33 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X092618W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 03 04:00:03 2018  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	193885	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	306812	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	299094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	165485	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	144126	52.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.08%	
35) Dibromofluoromethane	5.50	113	123239	47.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.78%	
50) Toluene-d8	8.71	98	443896	51.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.70%	
62) 4-Bromofluorobenzene	11.14	95	155945	50.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.14%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	11293	5.32	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	4051	1.58	ug/l	96
44) Trichloroethene	7.21	130	257772	94.41	ug/l	96
49) 1,4-Dioxane	7.76	88	1195	15.59	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_X\DATA\VX100818\  
 Data File : VX005124.D  
 Acq On : 08 Oct 2018 13:48  
 Operator : JC/MD  
 Sample : J5189-07  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE-105D1-20180927

Quant Time: Oct 09 12:53:33 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X092618W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 03 04:00:03 2018  
 Response via : Initial Calibration

