

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
 Data File : VX005140.D
 Acq On : 08 Oct 2018 21:49
 Operator : JC/MD
 Sample : VX1008WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1008WBL02

Quant Time: Oct 09 07:39:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	189358	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	297600	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	289115	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	160334	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	141455	52.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.58%	
35) Dibromofluoromethane	5.50	113	119906	47.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.08%	
50) Toluene-d8	8.72	98	422832	50.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.86%	
62) 4-Bromofluorobenzene	11.14	95	152240	50.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
Data File : VX005140.D
Acq On : 08 Oct 2018 21:49
Operator : JC/MD
Sample : VX1008WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1008WBL02

Quant Time: Oct 09 07:39:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

