

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
 Data File : VX005145.D
 Acq On : 09 Oct 2018 00:03
 Operator : JC/MD
 Sample : J5303-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB07-20181004

Quant Time: Oct 09 08:04:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	203683	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	307007	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	308844	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	173622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	158420	54.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.90%	
35) Dibromofluoromethane	5.50	113	134324	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
50) Toluene-d8	8.72	98	447472	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
62) 4-Bromofluorobenzene	11.14	95	161173	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
Target Compounds						
16) Acetone	2.45	43	12712	7.48	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
Data File : VX005145.D
Acq On : 09 Oct 2018 00:03
Operator : JC/MD
Sample : J5303-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB07-20181004

Quant Time: Oct 09 08:04:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

