

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
 Data File : VX005147.D
 Acq On : 09 Oct 2018 00:56
 Operator : JC/MD
 Sample : J5189-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-131D2-20180927DL

Quant Time: Oct 09 09:33:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	191410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	286317	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	283810	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	156661	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	144393	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
35) Dibromofluoromethane	5.51	113	127326	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
50) Toluene-d8	8.71	98	419678	52.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.04%	
62) 4-Bromofluorobenzene	11.14	95	148515	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	70806	33.81	ug/l	100
44) Trichloroethene	7.21	130	33254	13.05	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
Data File : VX005147.D
Acq On : 09 Oct 2018 00:56
Operator : JC/MD
Sample : J5189-04DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-131D2-20180927DL

Quant Time: Oct 09 09:33:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

