

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
 Data File : VX005149.D
 Acq On : 09 Oct 2018 01:50
 Operator : JC/MD
 Sample : J5189-06DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-131D3-20180927DL

Quant Time: Oct 09 08:14:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	196106	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	288185	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	284730	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	156503	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	153619	54.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.68%	
35) Dibromofluoromethane	5.51	113	127028	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
50) Toluene-d8	8.72	98	421526	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
62) 4-Bromofluorobenzene	11.14	95	146899	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.38	101	68617	31.98	ug/l	99
44) Trichloroethene	7.21	130	5228	2.04	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
Data File : VX005149.D
Acq On : 09 Oct 2018 01:50
Operator : JC/MD
Sample : J5189-06DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-131D3-20180927DL

Quant Time: Oct 09 08:14:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

