

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
 Data File : VX005150.D
 Acq On : 09 Oct 2018 02:16
 Operator : JC/MD
 Sample : J5189-10DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-101D-20180928DL

Quant Time: Oct 09 09:35:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	193430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	282019	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	277701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152025	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	141814	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
35) Dibromofluoromethane	5.51	113	128989	53.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.94%	
50) Toluene-d8	8.72	98	410868	51.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.42%	
62) 4-Bromofluorobenzene	11.14	95	143125	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.38	101	6930	3.27	ug/l	98
44) Trichloroethene	7.21	130	108315	43.16	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
Data File : VX005150.D
Acq On : 09 Oct 2018 02:16
Operator : JC/MD
Sample : J5189-10DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-101D-20180928DL

Quant Time: Oct 09 09:35:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

