

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
 Data File : VX005152.D
 Acq On : 09 Oct 2018 03:10
 Operator : JC/MD
 Sample : J5189-09DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-101D2-20180928DL

Quant Time: Oct 09 08:21:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	195592	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	294074	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	295703	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	163836	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	147758	52.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.76%	
35) Dibromofluoromethane	5.50	113	123385	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
50) Toluene-d8	8.71	98	431253	52.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.10%	
62) 4-Bromofluorobenzene	11.14	95	154785	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	2535	1.18	ug/l	95
44) Trichloroethene	7.21	130	114036	43.57	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
Data File : VX005152.D
Acq On : 09 Oct 2018 03:10
Operator : JC/MD
Sample : J5189-09DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-101D2-20180928DL

Quant Time: Oct 09 08:21:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

