

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
 Data File : VX005153.D
 Acq On : 09 Oct 2018 03:37
 Operator : JC/MD
 Sample : J5189-08DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-105D2-20180927DL

Quant Time: Oct 09 08:22:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	193781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	296817	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	296410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166907	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	149010	53.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.66%	
35) Dibromofluoromethane	5.51	113	129555	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
50) Toluene-d8	8.71	98	437499	52.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.62%	
62) 4-Bromofluorobenzene	11.14	95	155695	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Trichloroethene	7.21	130	76204	28.85	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX100818\
Data File : VX005153.D
Acq On : 09 Oct 2018 03:37
Operator : JC/MD
Sample : J5189-08DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-105D2-20180927DL

Quant Time: Oct 09 08:22:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

