

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100819\
 Data File : VX012892.D
 Acq On : 08 Oct 2019 16:17
 Operator : JC/SP
 Sample : K5122-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-GMZ5-ER1

Quant Time: Oct 09 03:10:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	179960	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	304034	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	254813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	96642	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115474	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
35) Dibromofluoromethane	5.49	113	89994	51.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.14%	
50) Toluene-d8	8.71	98	351432	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
62) 4-Bromofluorobenzene	11.13	95	117408	44.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.80%	
Target Compounds						
16) Acetone	2.43	43	10564	8.910	ug/l	97
17) Carbon Disulfide	2.56	76	1477	0.285	ug/l	100
25) 2-Butanone	4.67	43	14020	8.524	ug/l	98
29) Tetrahydrofuran	5.12	42	17391	17.182	ug/l	98
95) Naphthalene	13.83	128	2097	0.317	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100819\
Data File : VX012892.D
Acq On : 08 Oct 2019 16:17
Operator : JC/SP
Sample : K5122-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-GMZ5-ER1

Quant Time: Oct 09 03:10:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

