

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100819\
 Data File : VX012893.D
 Acq On : 08 Oct 2019 16:40
 Operator : JC/SP
 Sample : K5122-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-FMW3413

Quant Time: Oct 09 03:12:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178087	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	300817	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	248689	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	96564	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	117100	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
35) Dibromofluoromethane	5.49	113	89910	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
50) Toluene-d8	8.71	98	355853	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
62) 4-Bromofluorobenzene	11.14	95	116873	44.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.34%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	1039	0.467	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100819\
Data File : VX012893.D
Acq On : 08 Oct 2019 16:40
Operator : JC/SP
Sample : K5122-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-FMW3413

Quant Time: Oct 09 03:12:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

