

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100819\
 Data File : VX012897.D
 Acq On : 08 Oct 2019 18:14
 Operator : JC/SP
 Sample : K5122-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-CHY07

Quant Time: Oct 09 03:22:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	175081	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	305507	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	254775	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97005	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117630	53.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.50%	
35) Dibromofluoromethane	5.49	113	89352	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
50) Toluene-d8	8.71	98	351817	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	11.14	95	117935	44.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.76%	
Target Compounds						
16) Acetone	2.43	43	10226	8.866	ug/l #	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100819\
Data File : VX012897.D
Acq On : 08 Oct 2019 18:14
Operator : JC/SP
Sample : K5122-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-CHY07

Quant Time: Oct 09 03:22:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

