

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100819\
 Data File : VX012905.D
 Acq On : 08 Oct 2019 21:43
 Operator : JC/SP
 Sample : VX1008WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1008WBL02

Quant Time: Oct 09 04:07:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	172177	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	292952	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	252335	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	105398	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	111768	51.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.88%	
35) Dibromofluoromethane	5.48	113	86767	51.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.20%	
50) Toluene-d8	8.71	98	342808	52.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.54%	
62) 4-Bromofluorobenzene	11.13	95	121741	47.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.56%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100819\
Data File : VX012905.D
Acq On : 08 Oct 2019 21:43
Operator : JC/SP
Sample : VX1008WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1008WBL02

Quant Time: Oct 09 04:07:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

