

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100819\
 Data File : VX012908.D
 Acq On : 08 Oct 2019 23:39
 Operator : JC/SP
 Sample : K5232-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-GMZ1CSS-TB1

Quant Time: Oct 09 04:29:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178205	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	300312	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	252548	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97218	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	115795	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
35) Dibromofluoromethane	5.49	113	88959	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
50) Toluene-d8	8.71	98	355430	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
62) 4-Bromofluorobenzene	11.13	95	118201	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100819\
Data File : VX012908.D
Acq On : 08 Oct 2019 23:39
Operator : JC/SP
Sample : K5232-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-GMZ1CSS-TB1

Quant Time: Oct 09 04:29:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

