

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100819\
 Data File : VX012915.D
 Acq On : 09 Oct 2019 02:22
 Operator : JC/SP
 Sample : K5232-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-056MW04

Quant Time: Oct 09 07:35:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	181632	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	309181	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	266084	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	109021	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.05	65	118268	52.10	ug/l	0.00
Spiked Amount			50.000			
Recovery			=	104.20%		
35) Dibromofluoromethane	5.48	113	90990	51.28	ug/l	0.00
Spiked Amount			50.000			
Recovery			=	102.56%		
50) Toluene-d8	8.71	98	363204	52.48	ug/l	0.00
Spiked Amount			50.000			
Recovery			=	104.96%		
62) 4-Bromofluorobenzene	11.14	95	128992	47.97	ug/l	0.00
Spiked Amount			50.000			
Recovery			=	95.94%		

Target Compounds

						Qvalue
4) Vinyl Chloride	1.40	62	19906	9.572	ug/l	99
16) Acetone	2.43	43	4500	3.761	ug/l	93
21) trans-1,2-Dichloroethene	3.16	96	1329	0.676	ug/l #	84
27) cis-1,2-Dichloroethene	4.59	96	40517	17.867	ug/l	89
37) Ethyl Acetate	4.83	43	5585	1.837	ug/l #	80
40) Benzene	6.14	78	2687	0.335	ug/l #	88
44) Trichloroethene	7.21	130	1063	0.477	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100819\
 Data File : VX012915.D
 Acq On : 09 Oct 2019 02:22
 Operator : JC/SP
 Sample : K5232-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-056MW04

Quant Time: Oct 09 07:35:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

