

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100819\
 Data File : VX012919.D
 Acq On : 09 Oct 2019 03:55
 Operator : JC/SP
 Sample : K5154-23
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-102D

Quant Time: Oct 09 07:44:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	184939	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	312016	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	264939	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	107112	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	118797	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
35) Dibromofluoromethane	5.49	113	92819	51.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.66%	
50) Toluene-d8	8.71	98	364683	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
62) 4-Bromofluorobenzene	11.14	95	125913	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
Target Compounds						
16) Acetone	2.43	43	5195	4.264	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100819\
Data File : VX012919.D
Acq On : 09 Oct 2019 03:55
Operator : JC/SP
Sample : K5154-23
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-102D

Quant Time: Oct 09 07:44:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

