

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100820\
 Data File : VX018834.D
 Acq On : 08 Oct 2020 13:10
 Operator : JC/SP
 Sample : VX1008WBL01
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK82

Quant Time: Oct 09 05:04:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 09 04:43:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.82	114	125318	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.09	117	127242	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	57533	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	40188	5.49	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.80%
7) Chloroethane-d5	1.70	69	32520	5.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.60%
11) 1,1-Dichloroethene-d2	2.34	63	70364	4.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.20%
20) 2-Butanone-d5	4.55	46	100220	52.19	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.38%
24) Chloroform-d	5.14	84	85848	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
26) 1,2-Dichloroethane-d4	6.03	65	51095	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
32) Benzene-d6	6.04	84	155062	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	7.37	67	47332	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	8.70	98	130923	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.00%
45) trans-1,3-Dichloropropene-	8.99	79	20612	4.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.80%
48) 2-Hexanone-d5	9.43	63	77576	47.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.12%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	36473	4.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.20%
65) 1,2-Dichlorobenzene-d4	12.36	152	54622	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds					Ovalue
16) Methylene chloride	2.84	84	2940	0.388 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100820\
Data File : VX018834.D
Acq On : 08 Oct 2020 13:10
Operator : JC/SP
Sample : VX1008WBL01
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK82

Quant Time: Oct 09 05:04:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 09 04:43:33 2020
Response via : Initial Calibration

