

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX100820\
 Data File : VX018839.D
 Acq On : 08 Oct 2020 15:23
 Operator : JC/SP
 Sample : L4283-03
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFZR5

Quant Time: Oct 09 05:51:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 09 04:43:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	167550	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	160369	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	75815	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	31108	3.18	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.60%
7) Chloroethane-d5	1.70	69	30372	3.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.40%
11) 1,1-Dichloroethene-d2	2.34	63	56187	2.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	50.80%#
20) 2-Butanone-d5	4.54	46	136420	53.14	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.28%
24) Chloroform-d	5.14	84	93527	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.00%
26) 1,2-Dichloroethane-d4	6.03	65	54123	3.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.80%
32) Benzene-d6	6.04	84	162265	4.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.40%
36) 1,2-Dichloropropane-d6	7.37	67	53406	4.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.80%
41) Toluene-d8	8.70	98	140476	3.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.60%
45) trans-1,3-Dichloropropene-	8.99	79	22279	3.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.80%
48) 2-Hexanone-d5	9.43	63	107071	51.54	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.08%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	43389	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.60%
65) 1,2-Dichlorobenzene-d4	12.36	152	62898	4.50	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.00%

Target Compounds					Ovalue
3) Chloromethane	1.31	50	8486	1.047 ug/L	95
13) Acetone	2.45	43	15645	11.162 ug/L	98
14) Carbon disulfide	2.56	76	5817	0.223 ug/L #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100820\
Data File : VX018839.D
Acq On : 08 Oct 2020 15:23
Operator : JC/SP
Sample : L4283-03
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFZR5

Quant Time: Oct 09 05:51:08 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 09 04:43:33 2020
Response via : Initial Calibration

