

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100820\
 Data File : VX018842.D
 Acq On : 08 Oct 2020 16:39
 Operator : JC/SP
 Sample : L4283-05
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK01

Quant Time: Oct 09 06:32:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 09 04:43:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	134867	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	135496	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	62536	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	29210	3.71	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.20%
7) Chloroethane-d5	1.70	69	30120	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.80%
11) 1,1-Dichloroethene-d2	2.34	63	53893	3.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.60%
20) 2-Butanone-d5	4.56	46	131783	63.77	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.54%
24) Chloroform-d	5.14	84	92417	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
26) 1,2-Dichloroethane-d4	6.03	65	54647	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
32) Benzene-d6	6.05	84	156881	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
36) 1,2-Dichloropropane-d6	7.37	67	50814	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.40%
41) Toluene-d8	8.70	98	135775	4.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.80%
45) trans-1,3-Dichloropropene-	8.99	79	22135	4.22	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.40%
48) 2-Hexanone-d5	9.43	63	102935	58.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	117.28%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	42274	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.80%
65) 1,2-Dichlorobenzene-d4	12.36	152	60068	5.21	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.20%

Target Compounds					Ovalue
16) Methylene chloride	2.83	84	3002	0.369 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100820\
Data File : VX018842.D
Acq On : 08 Oct 2020 16:39
Operator : JC/SP
Sample : L4283-05
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK01

Quant Time: Oct 09 06:32:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 09 04:43:33 2020
Response via : Initial Calibration

