

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100820\
 Data File : VX018846.D
 Acq On : 08 Oct 2020 18:39
 Operator : JC/SP
 Sample : L4290-05
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK01

Quant Time: Oct 09 06:51:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 09 04:43:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	123339	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	126029	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	55124	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	27073	3.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.20%
7) Chloroethane-d5	1.70	69	28102	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.60%
11) 1,1-Dichloroethene-d2	2.34	63	52007	3.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.00%
20) 2-Butanone-d5	4.56	46	126768	67.08	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	134.16%#
24) Chloroform-d	5.14	84	86549	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	6.03	65	53690	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
32) Benzene-d6	6.04	84	147159	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
36) 1,2-Dichloropropane-d6	7.37	67	47559	5.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.00%
41) Toluene-d8	8.70	98	124273	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.40%
45) trans-1,3-Dichloropropene-	8.99	79	20426	4.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.80%
48) 2-Hexanone-d5	9.43	63	100983	61.85	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.70%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	40884	5.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.00%
65) 1,2-Dichlorobenzene-d4	12.36	152	55195	5.43	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100820\
Data File : VX018846.D
Acq On : 08 Oct 2020 18:39
Operator : JC/SP
Sample : L4290-05
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK01

Quant Time: Oct 09 06:51:48 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR100220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 09 04:43:33 2020
Response via : Initial Calibration

