

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024665.D
 Acq On : 08 Oct 2021 12:18
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 08 16:26:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	178056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	251320	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124754	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	131519	52.409	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.820%	
35) Dibromofluoromethane	5.391	113	101480	60.782	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 121.560%	
50) Toluene-d8	8.653	98	330875	53.845	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.700%	
62) 4-Bromofluorobenzene	11.085	95	135341	56.017	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.040%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	119388	66.635	ug/l	97
3) Chloromethane	1.295	50	93697	55.480	ug/l	99
4) Vinyl Chloride	1.374	62	99916	56.559	ug/l	97
5) Bromomethane	1.599	94	68025	57.212	ug/l	98
6) Chloroethane	1.679	64	52734	44.845	ug/l	100
7) Trichlorofluoromethane	1.886	101	148097	47.505	ug/l	98
8) Diethyl Ether	2.136	74	48453	45.662	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.331	101	88074	50.100	ug/l	95
10) Methyl Iodide	2.453	142	120306	48.910	ug/l	92
11) Tert butyl alcohol	2.977	59	120420	204.050	ug/l	98
12) 1,1-Dichloroethene	2.319	96	82156	50.334	ug/l	91
13) Acrolein	2.240	56	42338	309.448	ug/l	100
14) Allyl chloride	2.666	41	153965	51.595	ug/l	94
15) Acrylonitrile	3.069	53	267988	235.066	ug/l	98
16) Acetone	2.386	43	349714	333.685	ug/l	90
17) Carbon Disulfide	2.514	76	223563	53.120	ug/l	99
18) Methyl Acetate	2.709	43	177886	47.339	ug/l	93
19) Methyl tert-butyl Ether	3.123	73	317019	51.537	ug/l	98
20) Methylene Chloride	2.788	84	99737	53.191	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	95637	51.600	ug/l	96
22) Diisopropyl ether	3.770	45	302896	49.681	ug/l	97
23) Vinyl Acetate	3.727	43	1271018	257.447	ug/l	96
24) 1,1-Dichloroethane	3.611	63	172722	51.024	ug/l	99
25) 2-Butanone	4.562	43	472777	273.561	ug/l	95
26) 2,2-Dichloropropane	4.483	77	163518	53.699	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	110358	50.943	ug/l	95
28) Bromochloromethane	4.904	49	69190	45.025	ug/l	100
29) Tetrahydrofuran	5.013	42	256169	232.625	ug/l	92
30) Chloroform	5.099	83	188889	50.524	ug/l	99
31) Cyclohexane	5.477	56	155464	51.736	ug/l	100
32) 1,1,1-Trichloroethane	5.391	97	173710	52.336	ug/l	100
36) 1,1-Dichloropropene	5.696	75	140808	58.443	ug/l	99
37) Ethyl Acetate	4.721	43	154534	51.374	ug/l	96
38) Carbon Tetrachloride	5.684	117	154756	59.008	ug/l	100
39) Methylcyclohexane	7.385	83	160103	56.753	ug/l	99
40) Benzene	6.044	78	393573	57.128	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024665.D
 Acq On : 08 Oct 2021 12:18
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 08 16:26:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	84495	54.393	ug/l	92
42) 1,2-Dichloroethane	6.092	62	160127	55.345	ug/l	94
43) Isopropyl Acetate	6.349	43	247403	54.100	ug/l	94
44) Trichloroethene	7.129	130	104235	54.990	ug/l	97
45) 1,2-Dichloropropane	7.434	63	94417	53.121	ug/l	96
46) Dibromomethane	7.586	93	69423	52.088	ug/l	98
47) Bromodichloromethane	7.824	83	130195	53.301	ug/l	99
48) Methyl methacrylate	7.696	41	124283	54.886	ug/l	87
49) 1,4-Dioxane	7.659	88	46249	891.710	ug/l #	96
51) 4-Methyl-2-Pentanone	8.580	43	810714	268.535	ug/l	91
52) Toluene	8.720	92	232907	52.335	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	148312	53.570	ug/l	100
54) cis-1,3-Dichloropropene	8.373	75	153126	54.395	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	98300	53.419	ug/l	97
56) Ethyl methacrylate	9.122	69	136287	46.533	ug/l #	84
57) 1,3-Dichloropropane	9.311	76	157444	50.709	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.245	63	352579	243.881	ug/l	99
59) 2-Hexanone	9.433	43	653032	274.667	ug/l	88
60) Dibromochloromethane	9.525	129	103857	49.746	ug/l	99
61) 1,2-Dibromoethane	9.616	107	105403	52.176	ug/l	100
64) Tetrachloroethene	9.275	164	93768	59.087	ug/l	95
65) Chlorobenzene	10.080	112	259608	54.349	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	98612	55.393	ug/l	99
67) Ethyl Benzene	10.195	91	468674	55.428	ug/l	100
68) m/p-Xylenes	10.305	106	358056	110.428	ug/l	99
69) o-Xylene	10.647	106	171907	53.615	ug/l	100
70) Styrene	10.659	104	279477	52.127	ug/l	97
71) Bromoform	10.805	173	75797	48.911	ug/l #	99
73) Isopropylbenzene	10.964	105	472834	53.612	ug/l	100
74) N-amyl acetate	10.848	43	201261	49.612	ug/l	91
75) 1,1,2,2-Tetrachloroethane	11.214	83	152521	48.901	ug/l	100
76) 1,2,3-Trichloropropane	11.244	75	184062	64.969	ug/l	88
77) Bromobenzene	11.201	156	117883	51.944	ug/l	96
78) n-propylbenzene	11.305	91	545891	52.948	ug/l	99
79) 2-Chlorotoluene	11.366	91	322762	51.392	ug/l	99
80) 1,3,5-Trimethylbenzene	11.457	105	417695	55.439	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.025	75	51321	49.931	ug/l	92
82) 4-Chlorotoluene	11.457	91	385768	52.205	ug/l	100
83) tert-Butylbenzene	11.720	119	395923	52.416	ug/l	95
84) 1,2,4-Trimethylbenzene	11.756	105	420460	56.126	ug/l	98
85) sec-Butylbenzene	11.896	105	473209	50.922	ug/l	100
86) p-Isopropyltoluene	12.012	119	442085	55.790	ug/l	97
87) 1,3-Dichlorobenzene	11.976	146	219301	50.741	ug/l	100
88) 1,4-Dichlorobenzene	12.043	146	220863	49.470	ug/l	99
89) n-Butylbenzene	12.335	91	395863	56.419	ug/l	98
90) Hexachloroethane	12.543	117	69791	53.561	ug/l	96
91) 1,2-Dichlorobenzene	12.335	146	211553	49.513	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	37567	49.884	ug/l	91
93) 1,2,4-Trichlorobenzene	13.591	180	145981	53.663	ug/l	98
94) Hexachlorobutadiene	13.725	225	65098	52.313	ug/l	99
95) Naphthalene	13.780	128	482799	55.449	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	148989	53.448	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024665.D
Acq On : 08 Oct 2021 12:18
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 08 16:26:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024665.D
 Acq On : 08 Oct 2021 12:18
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :


Quant Time: Oct 08 16:26:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 733
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:168 Resp: 178056
Ion Ratio Lower Upper
168 100
99 59.3 45.8 68.6




#2
Dichlorodifluoromethane
Concen: 66.635 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 85 Resp: 119388
Ion Ratio Lower Upper
85 100
87 31.0 16.4 49.2







Abundance Scan 34 (1.294 min): VX024453.D\data.ms (-30)






#3
Chloromethane
Concen: 55.480 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18


Tgt Ion: 50 Resp: 93697
Ion Ratio Lower Upper
50 100
52 32.4 26.4 39.6



#4
Vinyl Chloride
Concen: 56.559 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 62 Resp: 99916
Ion Ratio Lower Upper
62 100
64 30.3 25.7 38.5









#5
Bromomethane
Concen: 57.212 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 68025
Ion Ratio Lower Upper
94 100
96 93.1 76.4 114.6



#6
Chloroethane
Concen: 44.845 ug/l
RT: 1.679 min Scan# 97
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 64 Resp: 52734
Ion Ratio Lower Upper
64 100
66 31.7 25.4 38.0









#7
Trichlorofluoromethane
Concen: 47.505 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 148097
Ion Ratio Lower Upper
101 100
103 64.8 53.2 79.8



#8
Diethyl Ether
Concen: 45.662 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 74 Resp: 48453
Ion Ratio Lower Upper
74 100
45 93.4 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.100 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.006 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:101 Resp: 88074
Ion Ratio Lower Upper
101 100
85 46.5 34.9 52.3
151 81.9 61.4 92.2



#10
Methyl Iodide
Concen: 48.910 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:142 Resp: 120306
Ion Ratio Lower Upper
142 100
127 48.3 33.3 49.9
141 14.3 11.4 17.0









#11

Tert butyl alcohol
Concen: 204.050 ug/l
RT: 2.977 min Scan# 310
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 59 Resp: 120420
Ion Ratio Lower Upper
59 100
57 10.5 7.8 11.8




#12

1,1-Dichloroethene
Concen: 50.334 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 96 Resp: 82156
Ion Ratio Lower Upper
96 100
61 168.9 124.6 187.0
98 67.5 50.6 76.0









#13
Acrolein
Concen: 309.448 ug/l
RT: 2.240 min Scan# 189
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 42338
Ion Ratio Lower Upper
56 100
55 70.8 56.4 84.6



#14
Allyl chloride
Concen: 51.595 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 41 Resp: 153965
Ion Ratio Lower Upper
41 100
39 68.9 51.6 77.4
76 33.3 30.5 45.7









#15
Acrylonitrile
Concen: 235.066 ug/l
RT: 3.069 min Scan# 325
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 267988
Ion Ratio Lower Upper
53 100
52 82.4 65.3 97.9
51 37.0 28.2 42.2



#16
Acetone
Concen: 333.685 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 43 Resp: 349714
Ion Ratio Lower Upper
43 100
58 28.9 27.8 41.8









#17
Carbon Disulfide
Concen: 53.120 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 223563
Ion Ratio Lower Upper
76 100
78 9.3 7.2 10.8



#18
Methyl Acetate
Concen: 47.339 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 43 Resp: 177886
Ion Ratio Lower Upper
43 100
74 23.9 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 51.537 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 73 Resp: 317019
Ion Ratio Lower Upper
73 100
57 20.3 17.1 25.7



#20
Methylene Chloride
Concen: 53.191 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 84 Resp: 99737
Ion Ratio Lower Upper
84 100
49 121.1 88.6 132.8
51 36.8 27.0 40.6
86 64.2 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 51.600 ug/l
RT: 3.093 min Scan# 329
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 96 Resp: 95637
Ion Ratio Lower Upper
96 100
61 141.0 107.7 161.5
98 63.8 50.4 75.6



#22
Diisopropyl ether
Concen: 49.681 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 45 Resp: 302896
Ion Ratio Lower Upper
45 100
43 70.0 58.0 87.0
87 29.8 25.4 38.0
59 10.9 10.2 15.2









#23
 Vinyl Acetate
 Concen: 257.447 ug/l
 RT: 3.727 min Scan# 433
 Delta R.T. -0.006 min
 Lab File: VX024665.D
 Acq: 08 Oct 2021 12:18




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 1271018
 Ion Ratio Lower Upper
 43 100
 86 10.7 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 51.024 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX024665.D
 Acq: 08 Oct 2021 12:18

Tgt Ion: 63 Resp: 172722
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.8 11.2
 100 4.4 2.5 7.5









#25

2-Butanone
Concen: 273.561 ug/l
RT: 4.562 min Scan# 570
Delta R.T. -0.006 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 43 Resp: 472777
Ion Ratio Lower Upper
43 100
72 26.9 23.7 35.5




#26

2,2-Dichloropropane
Concen: 53.699 ug/l
RT: 4.483 min Scan# 557
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 77 Resp: 163518
Ion Ratio Lower Upper
77 100
97 22.1 11.7 35.1









#27
 cis-1,2-Dichloroethene
 Concen: 50.943 ug/l
 RT: 4.495 min Scan# 559
 Delta R.T. 0.000 min
 Lab File: VX024665.D
 Acq: 08 Oct 2021 12:18

Instrument :
 MSVOA_X
 ClientSampled :

Tgt Ion: 96 Resp: 110358




Ion	Ratio	Lower	Upper
96	100		
61	150.8	0.0	285.6
98	64.3	0.0	129.6




#28
 Bromochloromethane
 Concen: 45.025 ug/l
 RT: 4.904 min Scan# 626
 Delta R.T. -0.006 min
 Lab File: VX024665.D
 Acq: 08 Oct 2021 12:18

Tgt Ion: 49 Resp: 69190

Ion	Ratio	Lower	Upper
49	100		
129	2.2	0.0	4.0
130	85.2	68.2	102.2





#29

Tetrahydrofuran

Concen: 232.625 ug/l

RT: 5.013 min Scan# 644





Delta R.T. -0.006 min

Lab File: VX024665.D

Acq: 08 Oct 2021 12:18

Tgt Ion: 42 Resp: 256169

Ion	Ratio	Lower	Upper
42	100		
72	45.9	41.7	62.5
71	42.7	38.4	57.6



#30

Chloroform

Concen: 50.524 ug/l

RT: 5.099 min Scan# 658




Delta R.T. -0.006 min


Lab File: VX024665.D

Acq: 08 Oct 2021 12:18

Tgt Ion: 83 Resp: 188889

Ion	Ratio	Lower	Upper
83	100		
85	64.2	52.0	78.0









#31
Cyclohexane
Concen: 51.736 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 56 Resp: 155464
Ion Ratio Lower Upper
56 100
69 32.0 25.3 37.9
84 88.9 71.4 107.0



#32
1,1,1-Trichloroethane
Concen: 52.336 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 97 Resp: 173710
Ion Ratio Lower Upper
97 100
99 64.5 51.5 77.3
61 43.1 34.2 51.4









#33
1,2-Dichloroethane-d4
Concen: 52.409 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 65 Resp: 131519
Ion Ratio Lower Upper
65 100
67 50.3 0.0 105.4



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.006 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 114 Resp: 251320
Ion Ratio Lower Upper
114 100
63 21.0 0.0 40.8
88 15.8 0.0 32.8









#35

Dibromofluoromethane
Concen: 60.782 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Instrument :
MSVOA_X
ClientSampleId :




Tgt Ion:113 Resp: 101480
Ion Ratio Lower Upper
113 100
111 101.5 81.9 122.9
192 19.6 16.7 25.1




#36

1,1-Dichloropropene
Concen: 58.443 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 75 Resp: 140808
Ion Ratio Lower Upper
75 100
110 36.1 18.1 54.4
77 30.8 25.3 37.9





#37

Ethyl Acetate


Concen: 51.374 ug/l

RT: 4.721 min Scan# 596




Delta R.T. -0.006 min

Lab File: VX024665.D

Acq: 08 Oct 2021 12:18



Tgt Ion:	43	Resp:	154534
Ion	Ratio	Lower	Upper
43	100		
61	13.9	12.8	19.2
70	11.5	10.5	15.7



#38

Carbon Tetrachloride


Concen: 59.008 ug/l

RT: 5.684 min Scan# 754



Delta R.T. -0.006 min


Lab File: VX024665.D

Acq: 08 Oct 2021 12:18



Tgt Ion:	117	Resp:	154756
Ion	Ratio	Lower	Upper
117	100		
119	95.3	76.2	114.2
121	31.2	24.8	37.2





#39

Methylcyclohexane

Concen: 56.753 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX024665.D

Acq: 08 Oct 2021 12:18





Tgt Ion: 83 Resp: 160103

Ion Ratio Lower Upper

83 100

55 74.4 58.9 88.3

98 47.2 38.1 57.1



#40

Benzene

Concen: 57.128 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX024665.D




Acq: 08 Oct 2021 12:18


Tgt Ion: 78 Resp: 393573

Ion Ratio Lower Upper

78 100

77 23.2 18.9 28.3









#41
Methacrylonitrile
Concen: 54.393 ug/l
RT: 4.928 min Scan# 630
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 41 Resp: 84495




Ion	Ratio	Lower	Upper
41	100		
39	56.3	42.5	63.7
67	70.2	65.0	97.6
52	29.2	25.3	37.9




#42
1,2-Dichloroethane
Concen: 55.345 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 62 Resp: 160127

Ion	Ratio	Lower	Upper
62	100		
98	8.0	0.0	20.2









#43
Isopropyl Acetate
Concen: 54.100 ug/l
RT: 6.349 min Scan# 863
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 43 Resp: 247403




Ion	Ratio	Lower	Upper
43	100		
61	19.3	18.2	27.4
87	13.7	12.5	18.7




#44
Trichloroethene
Concen: 54.990 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:130 Resp: 104235

Ion	Ratio	Lower	Upper
130	100		
95	96.4	0.0	198.0









#45
1,2-Dichloropropane
Concen: 53.121 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 94417
Ion Ratio Lower Upper
63 100
65 29.9 25.7 38.5



#46
Dibromomethane
Concen: 52.088 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 93 Resp: 69423
Ion Ratio Lower Upper
93 100
95 82.9 66.8 100.2
174 100.6 83.1 124.7









#47
Bromodichloromethane
Concen: 53.301 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 130195
Ion Ratio Lower Upper
83 100
85 64.8 51.4 77.0
127 8.4 6.8 10.2



#48
Methyl methacrylate
Concen: 54.886 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 41 Resp: 124283
Ion Ratio Lower Upper
41 100
69 79.3 76.0 114.0
39 61.8 44.6 67.0









#49
1,4-Dioxane
Concen: 891.710 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. -0.006 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 88 Resp: 46249
Ion Ratio Lower Upper
88 100
43 35.0 23.2 34.8#
58 68.1 55.0 82.4



#50
Toluene-d8
Concen: 53.845 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 98 Resp: 330875
Ion Ratio Lower Upper
98 100
100 63.7 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 268.535 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 810714
Ion Ratio Lower Upper
43 100
58 38.5 35.4 53.0



#52
Toluene
Concen: 52.335 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 92 Resp: 232907
Ion Ratio Lower Upper
92 100
91 174.8 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 53.570 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 75 Resp: 148312
Ion Ratio Lower Upper
75 100
77 31.3 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 54.395 ug/l
RT: 8.373 min Scan# 1195
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 75 Resp: 153126
Ion Ratio Lower Upper
75 100
77 30.2 25.7 38.5
39 52.4 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 53.419 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 98300




Ion	Ratio	Lower	Upper
97	100		
83	83.1	69.3	103.9
85	53.9	44.3	66.5
99	62.0	50.3	75.5



#56
Ethyl methacrylate
Concen: 46.533 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 69 Resp: 136287

Ion	Ratio	Lower	Upper
69	100		
41	69.1	46.8	70.2
39	41.0	24.7	37.1









#57
1,3-Dichloropropane
Concen: 50.709 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 157444
Ion Ratio Lower Upper
76 100
78 32.8 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 243.881 ug/l
RT: 8.245 min Scan# 1174
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 63 Resp: 352579
Ion Ratio Lower Upper
63 100
106 28.7 23.5 35.3









#59
2-Hexanone
Concen: 274.667 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 653032
Ion Ratio Lower Upper
43 100
58 52.9 30.9 92.8



#60
Dibromochloromethane
Concen: 49.746 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 129 Resp: 103857
Ion Ratio Lower Upper
129 100
127 77.6 38.4 115.2









#61
1,2-Dibromoethane
Concen: 52.176 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 105403
Ion Ratio Lower Upper
107 100
109 94.4 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 56.017 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 95 Resp: 135341
Ion Ratio Lower Upper
95 100
174 80.5 0.0 154.6
176 78.0 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 229293
Ion Ratio Lower Upper
117 100
82 58.6 47.4 71.2
119 32.4 26.6 39.8



#64
Tetrachloroethene
Concen: 59.087 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.006 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:164 Resp: 93768
Ion Ratio Lower Upper
164 100
166 126.9 103.4 155.2
129 98.5 70.8 106.2
131 93.1 72.5 108.7









#65
Chlorobenzene
Concen: 54.349 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:112 Resp: 259608
Ion Ratio Lower Upper
112 100
114 31.8 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 55.393 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:131 Resp: 98612
Ion Ratio Lower Upper
131 100
133 94.3 47.7 143.1
119 64.2 32.1 96.5









#67
Ethyl Benzene
Concen: 55.428 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 468674
Ion Ratio Lower Upper
91 100
106 30.8 24.6 37.0



#68
m/p-Xylenes
Concen: 110.428 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 106 Resp: 358056
Ion Ratio Lower Upper
106 100
91 205.3 162.8 244.2









#69
o-Xylene
Concen: 53.615 ug/l
RT: 10.647 min Scan# 1568
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:106 Resp: 171907
Ion Ratio Lower Upper
106 100
91 216.0 108.1 324.2



#70
Styrene
Concen: 52.127 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:104 Resp: 279477
Ion Ratio Lower Upper
104 100
78 53.1 40.0 60.0
103 55.7 43.4 65.0









#71
Bromoform
Concen: 48.911 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:173 Resp: 75797
Ion Ratio Lower Upper
173 100
175 48.8 23.9 71.8
254 0.1 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:152 Resp: 124754
Ion Ratio Lower Upper
152 100
115 88.2 41.1 123.3
150 174.8 0.0 349.0









#73
Isopropylbenzene
Concen: 53.612 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 472834
Ion Ratio Lower Upper
105 100
120 26.5 13.2 39.6



#74
N-ethyl acetate
Concen: 49.612 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 43 Resp: 201261
Ion Ratio Lower Upper
43 100
70 41.7 39.0 58.6
55 23.8 21.5 32.3
61 23.5 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 48.901 ug/l
RT: 11.214 min Scan# 1661
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 152521
Ion Ratio Lower Upper
83 100
131 10.5 5.1 15.2
85 63.8 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 64.969 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 75 Resp: 184062
Ion Ratio Lower Upper
75 100
77 31.2 19.1 57.5









#77
Bromobenzene
Concen: 51.944 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:156 Resp: 117883
Ion Ratio Lower Upper
156 100
77 156.9 82.8 248.6
158 97.8 49.1 147.4



#78
n-propylbenzene
Concen: 52.948 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 91 Resp: 545891
Ion Ratio Lower Upper
91 100
120 23.6 11.5 34.5









#79
2-Chlorotoluene
Concen: 51.392 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 91 Resp: 322762
Ion Ratio Lower Upper
91 100
126 33.9 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 55.439 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 105 Resp: 417695
Ion Ratio Lower Upper
105 100
120 47.4 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 49.931 ug/l
RT: 11.025 min Scan# 1630
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 51321
Ion Ratio Lower Upper
75 100
53 103.9 74.5 111.7
89 48.9 37.9 56.9



#82
4-Chlorotoluene
Concen: 52.205 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18


Tgt Ion: 91 Resp: 385768
Ion Ratio Lower Upper
91 100
126 29.7 14.9 44.5








#83
tert-Butylbenzene
Concen: 52.416 ug/l
RT: 11.720 min Scan# 1744
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Instrument :
MSVOA_X
ClientSampled :







Tgt Ion:119 Resp: 395923
Ion Ratio Lower Upper
119 100
91 60.7 27.9 83.7
134 23.0 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 56.126 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:105 Resp: 420460
Ion Ratio Lower Upper
105 100
120 43.9 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 50.922 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX024665.D
 Acq: 08 Oct 2021 12:18




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion:105 Resp: 473209
 Ion Ratio Lower Upper
 105 100
 134 20.0 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 55.790 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX024665.D
 Acq: 08 Oct 2021 12:18

Tgt Ion:119 Resp: 442085
 Ion Ratio Lower Upper
 119 100
 134 26.2 12.9 38.7
 91 25.5 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 50.741 ug/l
RT: 11.976 min Scan# 1786
Delta R.T. 0.007 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 219301
Ion Ratio Lower Upper
146 100
111 40.8 20.2 60.5
148 64.1 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 49.470 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:146 Resp: 220863
Ion Ratio Lower Upper
146 100
111 40.6 19.6 58.8
148 63.9 31.6 95.0









#89
n-Butylbenzene
Concen: 56.419 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 91 Resp: 395863
Ion Ratio Lower Upper
91 100
92 53.2 27.4 82.0
134 26.1 13.2 39.5



#90
Hexachloroethane
Concen: 53.561 ug/l
RT: 12.543 min Scan# 1879
Delta R.T. 0.001 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 117 Resp: 69791
Ion Ratio Lower Upper
117 100
201 77.5 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 49.513 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 211553
Ion Ratio Lower Upper
146 100
111 42.2 20.8 62.2
148 63.9 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.884 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion: 75 Resp: 37567
Ion Ratio Lower Upper
75 100
155 82.9 45.3 135.9
157 104.3 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 53.663 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	180	Resp:	145981
Ion Ratio	Lower	Upper	
180	100		
182	96.6	47.7	143.1
145	31.8	15.0	45.0



#94
Hexachlorobutadiene
Concen: 52.313 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:	225	Resp:	65098
Ion Ratio	Lower	Upper	
225	100		
223	62.8	31.1	93.2
227	63.7	31.6	95.0









#95
Naphthalene
Concen: 55.449 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:128 Resp: 482799
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 53.448 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX024665.D
Acq: 08 Oct 2021 12:18

Tgt Ion:180 Resp: 148989
Ion Ratio Lower Upper
180 100
182 94.0 47.6 142.9
145 33.6 16.1 48.3

