

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024671.D
 Acq On : 08 Oct 2021 15:36
 Operator : JC/MD
 Sample : M4130-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-RR-03-(SA)

Quant Time: Oct 08 16:31:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	144046	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247938	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105019	51.730	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.460%			
35) Dibromofluoromethane	5.397	113	82127	49.861	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.720%			
50) Toluene-d8	8.653	98	299795	49.453	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 98.900%			
62) 4-Bromofluorobenzene	11.085	95	117761	49.405	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.820%			
Target Compounds					Qvalue	
30) Chloroform	5.099	83	3813	1.261	ug/l #	81
40) Benzene	6.044	78	1551	0.228	ug/l	93
52) Toluene	8.726	92	2264	0.516	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024671.D
Acq On : 08 Oct 2021 15:36
Operator : JC/MD
Sample : M4130-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-RR-03-(SA)

Quant Time: Oct 08 16:31:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

