

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024684.D
 Acq On : 08 Oct 2021 20:41
 Operator : JC/MD
 Sample : M4156-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20-20211008-WC

Quant Time: Oct 11 05:51:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	140412	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	241010	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106646	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101954	51.520	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.040%
35) Dibromofluoromethane	5.397	113	79550	49.685	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	8.653	98	293882	49.871	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.740%
62) 4-Bromofluorobenzene	11.085	95	109899	47.432	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024684.D
Acq On : 08 Oct 2021 20:41
Operator : JC/MD
Sample : M4156-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20-20211008-WC

Quant Time: Oct 11 05:51:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

