

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024688.D
 Acq On : 08 Oct 2021 23:25
 Operator : JC/MD
 Sample : VX1008WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1008WBL02

Quant Time: Oct 11 05:52:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	233211	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93716	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100455	51.723	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.440%
35) Dibromofluoromethane	5.397	113	78483	50.658	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.320%
50) Toluene-d8	8.653	98	284388	49.874	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.740%
62) 4-Bromofluorobenzene	11.085	95	110821	49.430	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024688.D
Acq On : 08 Oct 2021 23:25
Operator : JC/MD
Sample : VX1008WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1008WBL02

Quant Time: Oct 11 05:52:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

