

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024691.D
 Acq On : 09 Oct 2021 00:35
 Operator : JC/MD
 Sample : M4096-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MWHR2-9-100521

Quant Time: Oct 11 06:37:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	150206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	259810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256217	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115419	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109643	51.792	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.580%
35) Dibromofluoromethane	5.397	113	86905	50.351	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.700%
50) Toluene-d8	8.653	98	318908	50.202	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.400%
62) 4-Bromofluorobenzene	11.085	95	120398	48.204	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024691.D
Acq On : 09 Oct 2021 00:35
Operator : JC/MD
Sample : M4096-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MWHR2-9-100521

Quant Time: Oct 11 06:37:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

