

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024693.D
 Acq On : 09 Oct 2021 01:22
 Operator : JC/MD
 Sample : M4096-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A3SEEP-100521

Quant Time: Oct 11 06:38:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	236329	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231350	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99874	51.803	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.600%	
35) Dibromofluoromethane	5.397	113	80925	51.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.080%	
50) Toluene-d8	8.653	98	289579	50.114	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.220%	
62) 4-Bromofluorobenzene	11.085	95	108119	47.588	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.180%	
Target Compounds						
16) Acetone	2.386	43	17393	11.715	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024693.D
Acq On : 09 Oct 2021 01:22
Operator : JC/MD
Sample : M4096-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A3SEEP-100521

Quant Time: Oct 11 06:38:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

