

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024698.D
 Acq On : 09 Oct 2021 03:19
 Operator : JC/MD
 Sample : PB139768TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139768TB

Quant Time: Oct 11 06:39:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	240420	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	244569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108260	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101331	52.479	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.960%	
35) Dibromofluoromethane	5.397	113	80593	50.460	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.920%	
50) Toluene-d8	8.653	98	303258	51.588	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.180%	
62) 4-Bromofluorobenzene	11.085	95	116206	50.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.560%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	16487	10.522	ug/l #	87
20) Methylene Chloride	2.794	84	5861	2.981	ug/l #	83
43) Isopropyl Acetate	6.354	43	17138	3.918	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024698.D
Acq On : 09 Oct 2021 03:19
Operator : JC/MD
Sample : PB139768TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139768TB

Quant Time: Oct 11 06:39:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

