

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024703.D
 Acq On : 09 Oct 2021 05:17
 Operator : JC/MD
 Sample : M4145-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-SO-DRUM-3-100721

Quant Time: Oct 11 06:40:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235644	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105576	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101581	53.419	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.840%	
35) Dibromofluoromethane	5.391	113	80026	51.120	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.240%	
50) Toluene-d8	8.653	98	299704	52.017	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.040%	
62) 4-Bromofluorobenzene	11.085	95	114553	50.567	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.140%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	18845	13.911	ug/l #	86
20) Methylene Chloride	2.794	84	5664	2.903	ug/l	94
43) Isopropyl Acetate	6.355	43	15469	3.608	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024703.D
Acq On : 09 Oct 2021 05:17
Operator : JC/MD
Sample : M4145-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-SO-DRUM-3-100721

Quant Time: Oct 11 06:40:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

