

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX100821\
 Data File : VX024687.D
 Acq On : 08 Oct 2021 22:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 11 05:52:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	50.000	58.630	-17.3	104	0.00
3 P	Chloromethane	50.000	47.743	4.5	88	0.00
4 C	Vinyl Chloride	50.000	51.560	-3.1#	93	0.00
5 T	Bromomethane	50.000	59.583	-19.2	107	0.00
6 T	Chloroethane	50.000	48.673	2.7	93	0.00
7 T	Trichlorofluoromethane	50.000	50.044	-0.1	91	0.00
8 T	Diethyl Ether	50.000	48.525	3.0	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.669	0.7	92	0.00
10 T	Methyl Iodide	50.000	37.579	24.8	68	0.00
11 T	Tert butyl alcohol	250.000	200.884	19.6	73	0.00
12 CM	1,1-Dichloroethene	50.000	50.706	-1.4#	95	0.00
13 T	Acrolein	250.000	285.783	-14.3	105	0.00
14 T	Allyl chloride	50.000	47.959	4.1	88	0.00
15 T	Acrylonitrile	250.000	232.497	7.0	87	0.00
16 T	Acetone	250.000	262.946	-5.2	92	0.00
17 T	Carbon Disulfide	50.000	48.078	3.8	92	0.00
18 T	Methyl Acetate	50.000	47.375	5.3	90	0.00
19 T	Methyl tert-butyl Ether	50.000	48.548	2.9	91	0.00
20 T	Methylene Chloride	50.000	50.721	-1.4	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.339	1.3	93	0.00
22 T	Diisopropyl ether	50.000	48.490	3.0	90	0.00
23 T	Vinyl Acetate	250.000	249.617	0.2	89	0.00
24 P	1,1-Dichloroethane	50.000	48.912	2.2	90	0.00
25 T	2-Butanone	250.000	257.003	-2.8	93	0.00
26 T	2,2-Dichloropropane	50.000	43.066	13.9	80	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.083	1.8	92	0.00
28 T	Bromochloromethane	50.000	49.082	1.8	100	0.00
29 T	Tetrahydrofuran	250.000	232.035	7.2	83	0.00
30 C	Chloroform	50.000	48.952	2.1#	91	0.00
31 T	Cyclohexane	50.000	48.986	2.0	90	0.00
32 T	1,1,1-Trichloroethane	50.000	50.307	-0.6	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.133	-4.3	98	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	93	0.00
35 S	Dibromofluoromethane	50.000	53.750	-7.5	100	0.00
36 T	1,1-Dichloropropene	50.000	49.589	0.8	92	0.00
37 T	Ethyl Acetate	50.000	46.116	7.8	87	0.00
38 T	Carbon Tetrachloride	50.000	50.784	-1.6	90	0.00
39 T	Methylcyclohexane	50.000	50.365	-0.7	89	0.00
40 TM	Benzene	50.000	49.732	0.5	93	0.00
41 T	Methacrylonitrile	50.000	47.756	4.5	87	0.00
42 TM	1,2-Dichloroethane	50.000	50.020	-0.0	91	0.00
43 T	Isopropyl Acetate	50.000	48.523	3.0	90	0.00
44 TM	Trichloroethene	50.000	50.874	-1.7	95	0.00
45 C	1,2-Dichloropropane	50.000	49.915	0.2#	92	0.00
46 T	Dibromomethane	50.000	49.019	2.0	92	0.00
47 T	Bromodichloromethane	50.000	51.747	-3.5	92	0.00
48 T	Methyl methacrylate	50.000	53.808	-7.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024687.D
 Acq On : 08 Oct 2021 22:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 11 05:52:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	873.446	12.7	76	0.00
50 S	Toluene-d8	50.000	52.865	-5.7	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	270.989	-8.4	97	0.00
52 CM	Toluene	50.000	51.177	-2.4#	93	0.00
53 T	t-1,3-Dichloropropene	50.000	49.651	0.7	89	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.969	-1.9	91	0.00
55 T	1,1,2-Trichloroethane	50.000	52.794	-5.6	96	0.00
56 T	Ethyl methacrylate	50.000	44.764	10.5	80	0.00
57 T	1,3-Dichloropropane	50.000	49.882	0.2	92	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	252.115	-0.8	88	0.00
59 T	2-Hexanone	250.000	270.285	-8.1	95	0.00
60 T	Dibromochloromethane	50.000	48.379	3.2	93	0.00
61 T	1,2-Dibromoethane	50.000	51.733	-3.5	93	0.00
62 S	4-Bromofluorobenzene	50.000	54.152	-8.3	100	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	56.569	-13.1	105	0.00
65 PM	Chlorobenzene	50.000	49.934	0.1	92	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.641	-3.3	94	0.00
67 C	Ethyl Benzene	50.000	51.194	-2.4#	92	0.00
68 T	m/p-Xylenes	100.000	102.844	-2.8	91	0.00
69 T	o-Xylene	50.000	49.848	0.3	90	0.00
70 T	Styrene	50.000	49.009	2.0	87	0.00
71 P	Bromoform	50.000	44.677	10.6	90	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	91	0.00
73 T	Isopropylbenzene	50.000	51.750	-3.5	92	0.00
74 T	N-ethyl acetate	50.000	48.659	2.7	87	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.715	2.6	89	0.00
76 T	1,2,3-Trichloropropane	50.000	49.383	1.2	90	0.00
77 T	Bromobenzene	50.000	50.635	-1.3	93	0.00
78 T	n-propylbenzene	50.000	50.426	-0.9	89	0.00
79 T	2-Chlorotoluene	50.000	50.338	-0.7	90	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.688	-5.4	91	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.355	11.3	82	0.00
82 T	4-Chlorotoluene	50.000	50.291	-0.6	90	0.00
83 T	tert-Butylbenzene	50.000	50.460	-0.9	88	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.287	-8.6	93	0.00
85 T	sec-Butylbenzene	50.000	48.516	3.0	84	0.00
86 T	p-Isopropyltoluene	50.000	52.498	-5.0	90	0.00
87 T	1,3-Dichlorobenzene	50.000	48.386	3.2	89	0.00
88 T	1,4-Dichlorobenzene	50.000	46.974	6.1	88	0.00
89 T	n-Butylbenzene	50.000	50.969	-1.9	88	0.00
90 T	Hexachloroethane	50.000	49.189	1.6	87	0.00
91 T	1,2-Dichlorobenzene	50.000	48.805	2.4	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.211	1.6	86	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.202	-0.4	90	0.00
94 T	Hexachlorobutadiene	50.000	45.857	8.3	83	0.00
95 T	Naphthalene	50.000	53.853	-7.7	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024687.D
Acq On : 08 Oct 2021 22:15
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 11 05:52:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.289	-0.6	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6