

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024708.D
 Acq On : 09 Oct 2021 07:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 11 06:42:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.503	0.626	-24.5	105	0.00
3 P	Chloromethane	0.474	0.494	-4.2	90	0.00
4 C	Vinyl Chloride	0.496	0.540	-8.9#	93	0.00
5 T	Bromomethane	0.334	0.356	-6.6	91	0.00
6 T	Chloroethane	0.378	0.324	14.3	88	0.00
7 T	Trichlorofluoromethane	0.875	0.900	-2.9	89	0.00
8 T	Diethyl Ether	0.298	0.304	-2.0	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.503	-1.8	89	0.00
10 T	Methyl Iodide	0.691	0.638	7.7	79	0.00
11 T	Tert butyl alcohol	0.178	0.139	21.9	72	0.00
12 CM	1,1-Dichloroethene	0.458	0.490	-7.0#	95	0.00
13 T	Acrolein	0.038	0.047	-23.7	107	0.00
14 T	Allyl chloride	0.838	0.843	-0.6	87	0.00
15 T	Acrylonitrile	0.320	0.311	2.8	86	0.00
16 T	Acetone	0.329	0.336	-2.1	94	0.00
17 T	Carbon Disulfide	1.230	1.222	0.7	93	0.00
18 T	Methyl Acetate	1.055	1.074	-1.8	92	0.00
19 T	Methyl tert-butyl Ether	1.727	1.783	-3.2	91	0.00
20 T	Methylene Chloride	0.599	0.571	4.7	96	0.00
21 T	trans-1,2-Dichloroethene	0.520	0.552	-6.2	95	0.00
22 T	Diisopropyl ether	1.712	1.760	-2.8	90	0.00
23 T	Vinyl Acetate	1.386	1.474	-6.3	90	0.00
24 P	1,1-Dichloroethane	0.951	0.999	-5.0	92	0.00
25 T	2-Butanone	0.485	0.530	-9.3	94	0.00
26 T	2,2-Dichloropropane	0.855	0.572	33.1#	59	0.00
27 T	cis-1,2-Dichloroethene	0.608	0.632	-3.9	93	0.00
28 T	Bromochloromethane	0.432	0.454	-5.1	101	0.00
29 T	Tetrahydrofuran	0.309	0.298	3.6	82	0.00
30 C	Chloroform	1.050	1.102	-5.0#	93	0.00
31 T	Cyclohexane	0.844	0.861	-2.0	88	0.00
32 T	1,1,1-Trichloroethane	0.932	1.000	-7.3	92	0.00
33 S	1,2-Dichloroethane-d4	0.705	0.740	-5.0	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.332	0.347	-4.5	95	0.00
36 T	1,1-Dichloropropene	0.479	0.485	-1.3	92	0.00
37 T	Ethyl Acetate	0.598	0.559	6.5	87	0.00
38 T	Carbon Tetrachloride	0.522	0.541	-3.6	90	0.00
39 T	Methylcyclohexane	0.561	0.561	0.0	87	0.00
40 TM	Benzene	1.371	1.408	-2.7	94	0.00
41 T	Methacrylonitrile	0.309	0.302	2.3	87	0.00
42 TM	1,2-Dichloroethane	0.576	0.593	-3.0	93	0.00
43 T	Isopropyl Acetate	0.910	0.900	1.1	90	0.00
44 TM	Trichloroethene	0.377	0.394	-4.5	96	0.00
45 C	1,2-Dichloropropane	0.354	0.368	-4.0#	95	0.00
46 T	Dibromomethane	0.265	0.265	0.0	92	0.00
47 T	Bromodichloromethane	0.486	0.512	-5.3	92	0.00
48 T	Methyl methacrylate	0.451	0.490	-8.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
 Data File : VX024708.D
 Acq On : 09 Oct 2021 07:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 11 06:42:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	74	0.00
50 S	Toluene-d8	1.223	1.255	-2.6	93	0.00
51 T	4-Methyl-2-Pentanone	0.601	0.661	-10.0	97	0.00
52 CM	Toluene	0.885	0.919	-3.8#	93	0.00
53 T	t-1,3-Dichloropropene	0.551	0.536	2.7	85	0.00
54 T	cis-1,3-Dichloropropene	0.560	0.564	-0.7	88	0.00
55 T	1,1,2-Trichloroethane	0.366	0.380	-3.8	93	0.00
56 T	Ethyl methacrylate	0.583	0.543	6.9	82	0.00
57 T	1,3-Dichloropropane	0.618	0.632	-2.3	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.288	0.304	-5.6	91	0.00
59 T	2-Hexanone	0.473	0.516	-9.1	94	0.00
60 T	Dibromochloromethane	0.371	0.409	-10.2	93	0.00
61 T	1,2-Dibromoethane	0.402	0.419	-4.2	93	0.00
62 S	4-Bromofluorobenzene	0.481	0.500	-4.0	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.346	0.398	-15.0	104	0.00
65 PM	Chlorobenzene	1.042	1.067	-2.4	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.409	-5.4	94	0.00
67 C	Ethyl Benzene	1.844	1.939	-5.2#	92	0.00
68 T	m/p-Xylenes	0.707	0.740	-4.7	91	0.00
69 T	o-Xylene	0.699	0.703	-0.6	89	0.00
70 T	Styrene	1.169	1.169	0.0	87	0.00
71 P	Bromoform	0.296	0.307	-3.7	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.535	3.730	-5.5	92	0.00
74 T	N-amyl acetate	1.626	1.635	-0.6	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.250	1.225	2.0	88	0.00
76 T	1,2,3-Trichloropropane	1.135	1.131	0.4	89	0.00
77 T	Bromobenzene	0.910	0.939	-3.2	94	0.00
78 T	n-propylbenzene	4.132	4.165	-0.8	88	0.00
79 T	2-Chlorotoluene	2.517	2.522	-0.2	88	0.00
80 T	1,3,5-Trimethylbenzene	3.020	3.220	-6.6	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.382	0.335	12.3	74	0.00
82 T	4-Chlorotoluene	2.962	3.022	-2.0	89	0.00
83 T	tert-Butylbenzene	3.027	3.080	-1.8	87	0.00
84 T	1,2,4-Trimethylbenzene	3.002	3.294	-9.7	93	0.00
85 T	sec-Butylbenzene	3.724	3.613	3.0	82	0.00
86 T	p-Isopropyltoluene	3.176	3.341	-5.2	89	0.00
87 T	1,3-Dichlorobenzene	1.732	1.697	2.0	88	0.00
88 T	1,4-Dichlorobenzene	1.789	1.696	5.2	87	0.00
89 T	n-Butylbenzene	2.812	2.878	-2.3	87	0.00
90 T	Hexachloroethane	0.522	0.529	-1.3	88	0.00
91 T	1,2-Dichlorobenzene	1.712	1.692	1.2	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.299	1.0	85	0.00
93 T	1,2,4-Trichlorobenzene	1.090	1.090	0.0	88	0.00
94 T	Hexachlorobutadiene	0.499	0.455	8.8	81	0.00
95 T	Naphthalene	3.490	3.786	-8.5	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100821\
Data File : VX024708.D
Acq On : 09 Oct 2021 07:13
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 11 06:42:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.117	1.121	-0.4	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6