

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100821\  
 Data File : VX024695.D  
 Acq On : 09 Oct 2021 02:09  
 Operator : JC/MD  
 Sample : M4096-08  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 35 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TRIP-BLANK

Quant Time: Oct 11 06:38:35 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092321W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 23 18:24:34 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.562  | 168   | 132345   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 230418   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 225779   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 99556    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.964  | 65    | 97921    | 52.498   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 105.000% |          |
| 35) Dibromofluoromethane    | 5.391  | 113   | 77692    | 50.755   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 101.520% |          |
| 50) Toluene-d8              | 8.653  | 98    | 281988   | 50.052   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 100.100% |          |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 106589   | 48.118   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 96.240%  |          |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100821\  
Data File : VX024695.D  
Acq On : 09 Oct 2021 02:09  
Operator : JC/MD  
Sample : M4096-08  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 35 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TRIP-BLANK

Quant Time: Oct 11 06:38:35 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092321W.M  
Quant Title : SW846 8260  
QLast Update : Thu Sep 23 18:24:34 2021  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.562 min Scan# 734  
Delta R.T. 0.006 min  
Lab File: VX024695.D  
Acq: 09 Oct 2021 02:09




Instrument :  
MSVOA\_X  
ClientSampleId :  
TRIP-BLANK


Tgt Ion:168 Resp: 132345  
Ion Ratio Lower Upper  
168 100  
99 57.9 45.8 68.6



#33  
1,2-Dichloroethane-d4  
Concen: 52.498 ug/l  
RT: 5.964 min Scan# 800  
Delta R.T. 0.000 min  
Lab File: VX024695.D  
Acq: 09 Oct 2021 02:09

Tgt Ion: 65 Resp: 97921  
Ion Ratio Lower Upper  
65 100  
67 49.5 0.0 105.4









#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.769 min Scan# 932  
Delta R.T. 0.000 min  
Lab File: VX024695.D  
Acq: 09 Oct 2021 02:09




Instrument :  
MSVOA\_X  
ClientSampleId :  
TRIP-BLANK


Tgt Ion:114 Resp: 230418  
Ion Ratio Lower Upper  
114 100  
63 20.3 0.0 40.8  
88 15.2 0.0 32.8



#35  
Dibromofluoromethane  
Concen: 50.755 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. -0.006 min  
Lab File: VX024695.D  
Acq: 09 Oct 2021 02:09

Tgt Ion:113 Resp: 77692  
Ion Ratio Lower Upper  
113 100  
111 102.7 81.9 122.9  
192 19.9 16.7 25.1









#50  
Toluene-d8  
Concen: 50.052 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. -0.000 min  
Lab File: VX024695.D  
Acq: 09 Oct 2021 02:09




Instrument :  
MSVOA\_X  
ClientSampleId :  
TRIP-BLANK


Tgt Ion: 98 Resp: 281988  
Ion Ratio Lower Upper  
98 100  
100 63.9 51.9 77.9



#62  
4-Bromofluorobenzene  
Concen: 48.118 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.000 min  
Lab File: VX024695.D  
Acq: 09 Oct 2021 02:09

Tgt Ion: 95 Resp: 106589  
Ion Ratio Lower Upper  
95 100  
174 81.9 0.0 154.6  
176 78.6 0.0 155.8









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX024695.D  
Acq: 09 Oct 2021 02:09

Instrument :  
MSVOA\_X  
ClientSampleId :  
TRIP-BLANK

Tgt Ion:117 Resp: 225779  
Ion Ratio Lower Upper  
117 100  
82 57.6 47.4 71.2  
119 32.8 26.6 39.8



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX024695.D  
Acq: 09 Oct 2021 02:09

Tgt Ion:152 Resp: 99556  
Ion Ratio Lower Upper  
152 100  
115 60.0 41.1 123.3  
150 156.4 0.0 349.0

