

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100824\
 Data File : VX043316.D
 Acq On : 08 Oct 2024 11:17
 Operator : JC/MD
 Sample : P4341-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 10324-0

Quant Time: Oct 09 01:34:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	77731	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	145038	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53303	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65838	51.863	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.720%
35) Dibromofluoromethane	5.385	113	47609	47.598	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.200%
50) Toluene-d8	8.647	98	172632	49.803	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.600%
62) 4-Bromofluorobenzene	11.079	95	60967	48.414	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.820%
Target Compounds						
16) Acetone	2.392	43	11057	20.857	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100824\
Data File : VX043316.D
Acq On : 08 Oct 2024 11:17
Operator : JC/MD
Sample : P4341-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
10324-0

Quant Time: Oct 09 01:34:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

