

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100824\
 Data File : VX043323.D
 Acq On : 08 Oct 2024 14:39
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 09 01:37:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	81286	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.751	114	152999	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	139394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60243	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69980	52.715	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.420%
35) Dibromofluoromethane	5.379	113	49938	47.328	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.660%
50) Toluene-d8	8.647	98	186963	51.131	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.260%
62) 4-Bromofluorobenzene	11.079	95	68474	51.546	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100824\
Data File : VX043323.D
Acq On : 08 Oct 2024 14:39
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 09 01:37:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

