

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100825\
 Data File : VX048068.D
 Acq On : 08 Oct 2025 12:22
 Operator : JC/MD
 Sample : Q3289-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-261

Quant Time: Oct 08 12:40:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	146300	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	311905	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	326457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	168435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	132704	56.986	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	113.980%	
35) Dibromofluoromethane	5.373	113	103341	49.403	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.800%	
50) Toluene-d8	8.635	98	350689	48.451	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	96.900%	
62) 4-Bromofluorobenzene	11.061	95	154980	56.418	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	112.840%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	21543	21.276	ug/l	99
20) Methylene Chloride	2.788	84	6210	2.897	ug/l #	82
43) Isopropyl Acetate	6.330	43	13181	2.456	ug/l	96
95) Naphthalene	13.756	128	160343	14.517	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100825\
Data File : VX048068.D
Acq On : 08 Oct 2025 12:22
Operator : JC/MD
Sample : Q3289-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-261

Quant Time: Oct 08 12:40:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

