

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100825\
 Data File : VX048069.D
 Acq On : 08 Oct 2025 12:43
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 09 04:53:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	148423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.744	114	300849	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	311637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	153321	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	131779	55.779	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.560%
35) Dibromofluoromethane	5.373	113	109204	54.124	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.240%
50) Toluene-d8	8.634	98	374947	53.706	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.420%
62) 4-Bromofluorobenzene	11.067	95	158963	59.995	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	119.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100825\
Data File : VX048069.D
Acq On : 08 Oct 2025 12:43
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 09 04:53:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

