

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005167.D
 Acq On : 09 Oct 2018 18:56
 Operator : JC/MD
 Sample : J5284-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-125D3-20181001

Quant Time: Oct 10 13:52:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	176796	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	278506	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	277121	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	155013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	136117	53.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.80%	
35) Dibromofluoromethane	5.51	113	115239	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
50) Toluene-d8	8.72	98	408697	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
62) 4-Bromofluorobenzene	11.14	95	145033	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	53599	27.709	ug/l	100
16) Acetone	2.45	43	6012	4.077	ug/l	92
27) cis-1,2-Dichloroethene	4.60	96	4591	1.960	ug/l	97
44) Trichloroethene	7.21	130	320267	129.218	ug/l	96
64) Tetrachloroethene	9.34	164	1859	0.682	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005167.D
Acq On : 09 Oct 2018 18:56
Operator : JC/MD
Sample : J5284-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-125D3-20181001

Quant Time: Oct 10 13:52:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

