

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005168.D
 Acq On : 09 Oct 2018 19:22
 Operator : JC/MD
 Sample : J5284-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-120D3-20181002

Quant Time: Oct 10 13:53:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	171794	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	276687	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	275086	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151841	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	134133	54.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.32%	
35) Dibromofluoromethane	5.50	113	113672	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
50) Toluene-d8	8.72	98	405927	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
62) 4-Bromofluorobenzene	11.14	95	143453	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	1978	1.052	ug/l	93
16) Acetone	2.45	43	7120	4.970	ug/l	94
44) Trichloroethene	7.21	130	86253	35.029	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005168.D
Acq On : 09 Oct 2018 19:22
Operator : JC/MD
Sample : J5284-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-120D3-20181002

Quant Time: Oct 10 13:53:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

