

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005171.D
 Acq On : 09 Oct 2018 20:43
 Operator : JC/MD
 Sample : J5284-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-103D3-20181003

Quant Time: Oct 10 07:28:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	175427	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	277912	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	277453	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152707	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	136504	54.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.94%	
35) Dibromofluoromethane	5.50	113	116513	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
50) Toluene-d8	8.72	98	413550	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
62) 4-Bromofluorobenzene	11.14	95	145031	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	4659	2.427	ug/l	96
16) Acetone	2.45	43	8054	5.505	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	1877	0.807	ug/l	87
30) Chloroform	5.21	83	2240	0.569	ug/l	88
44) Trichloroethene	7.21	130	1191943	481.939	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005171.D
Acq On : 09 Oct 2018 20:43
Operator : JC/MD
Sample : J5284-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-103D3-20181003

Quant Time: Oct 10 07:28:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

