

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005176.D
 Acq On : 09 Oct 2018 22:57
 Operator : JC/MD
 Sample : J5284-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-104D3-20181003

Quant Time: Oct 10 07:59:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	172996	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	273764	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	275502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	153517	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	138882	56.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.40%	
35) Dibromofluoromethane	5.50	113	114426	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
50) Toluene-d8	8.72	98	402733	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
62) 4-Bromofluorobenzene	11.14	95	145155	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
Target Compounds						
16) Acetone	2.45	43	3920	2.717	ug/l #	83
44) Trichloroethene	7.22	130	2780	1.141	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005176.D
Acq On : 09 Oct 2018 22:57
Operator : JC/MD
Sample : J5284-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-104D3-20181003

Quant Time: Oct 10 07:59:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

