

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005187.D
 Acq On : 10 Oct 2018 04:45
 Operator : JC/MD
 Sample : J5284-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-DUP03-20181003

Quant Time: Oct 10 16:27:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	186038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	284115	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	287118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	162840	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	140823	52.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.98%	
35) Dibromofluoromethane	5.49	113	121118	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
50) Toluene-d8	8.72	98	413582	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
62) 4-Bromofluorobenzene	11.14	95	152119	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
Target Compounds						
16) Acetone	2.45	43	5843	3.766	ug/l	95
44) Trichloroethene	7.21	130	2065	0.817	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005187.D
Acq On : 10 Oct 2018 04:45
Operator : JC/MD
Sample : J5284-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-DUP03-20181003

Quant Time: Oct 10 16:27:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

