

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005192.D
 Acq On : 10 Oct 2018 06:59
 Operator : JC/MD
 Sample : J5303-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW514-2024

Quant Time: Oct 10 08:49:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	182974	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	283332	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	282909	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	154432	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	143242	54.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.60%	
35) Dibromofluoromethane	5.49	113	128855	53.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.32%	
50) Toluene-d8	8.72	98	413227	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
62) 4-Bromofluorobenzene	11.14	95	147595	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
Target Compounds						
16) Acetone	2.45	43	3727	2.442	ug/l #	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005192.D
Acq On : 10 Oct 2018 06:59
Operator : JC/MD
Sample : J5303-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW514-2024

Quant Time: Oct 10 08:49:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

