

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005196.D
 Acq On : 10 Oct 2018 08:46
 Operator : JC/MD
 Sample : J5303-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW515-2024

Quant Time: Oct 10 14:07:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	185162	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	290092	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	292225	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	164016	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	144388	54.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.18%	
35) Dibromofluoromethane	5.51	113	120217	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
50) Toluene-d8	8.72	98	423487	51.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.62%	
62) 4-Bromofluorobenzene	11.14	95	153915	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
Target Compounds						
16) Acetone	2.45	43	3775	2.445	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005196.D
Acq On : 10 Oct 2018 08:46
Operator : JC/MD
Sample : J5303-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW515-2024

Quant Time: Oct 10 14:07:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

